



**BOARD OF DIRECTORS**

Lisa M. Borba, AICP  
PRESIDENT

Connstance Holdaway  
VICE PRESIDENT

Ernesto A. Avila, P.E.

Bette Boatman

John A. Burgh

**GENERAL MANAGER**

Jerry Brown

**Written Questions to  
Contra Costa Water District  
Request for Proposal (#2016) Addendum #1 dated November 6, 2019  
Questions and Answers for Procurement, Installation and Training of a Gas  
Chromatograph/Mass Spectrometer (GM/MS) and Solid Phase Micro Extraction Autosampler  
with Purge and Trap Sample Concentrator and Autosampler for Aqueous Samples**

1. I didn't see the Non-Collusion Affidavit mentioned in Attachment E (1.16 Method of Award f.) Perhaps I missed it but would you mind sending?

**Document is posted to the District's website and is due by November 13, 2019 not later than 11:00 A.M.**

2. Under the Specification Proposer Response Table, Section 1.1), Specification: "Transfer line must have a settable temperature limit of up to 450C for ideal transfer of components from GC to MS..."  
\*Which transfer line is in question? This sounds like the interface between the GC and MS?

**Yes the transfer line that carries analytes from the GC to the MS.**

3. Under the Specification Proposer Response Table, Section 1.1), Specification: "Backflush system to avoid carry over and contaminant removal." \*Can you please specify which part of the instrument you are wanting to backflush? Both GC columns? One GC column?

**Depending on your instrument setup answer the question accordingly. How does your system backflush to avoid carryover and contaminant removal?**

4. On the Pricing Sheet, Detail: "IT support for implementation and connection to CCWS LIMS." \* What are the specifications for connection to the LIMS system? Meaning, what are the columns of data to be transferred? In what format? ASCII?

**We are looking to find out if you provide an export for a LIMS system. You can list the formats offered. Would support be provided to get desired fields into the export? Fields may include sample site, parameter, reporting limit, dilution, and date of analysis. Other fields may also be requested.**

Thank you in advance for your cooperation.

Brian K. Jackson  
Purchasing Officer  
Contra Costa Water District

I acknowledge receipt of the foregoing Addendum #1.

Date: \_\_\_\_\_ Signature: \_\_\_\_\_